
CLIENT: Ross Environmental Associates
Project: 28-037 Hubbard Store
Lab Order: 1006003
Date Received: 6/2/2010

Work Order Sample Summary

Lab Sample ID	Client Sample ID	Collection Date	Collection Time
1006003-01A	MW-1	6/1/2010	9:53 AM

DATA COMMENT PAGE

Organic Data Qualifiers

ND	Indicates compound was analyzed for, but not detected at or above the reporting limit.
J	Indicates an estimated value. This flag is used either when estimating a concentration for tentatively identified compounds where a 1:1 response is assumed, or when the data indicates the presence of a compound that meets the identification criteria but the result is less than the sample quantitation limit but greater than the method detection limit.
H	Method prescribed holding time exceeded.
E	This flag identifies compounds whose concentrations exceed the calibration range of the instrument for that specific analysis.
B	This flag is used when the analyte is found in the associated blank as well as in the sample.
R	RPD outside accepted recovery limits
RL	Reporting limit; defined as the lowest concentration the laboratory can accurately quantitate.
S	Spike Recovery outside accepted recovery limits.
#	See Case Narrative

Micro Data Qualifiers

TNTC	Too numerous to count
------	-----------------------

Inorganic Data Qualifiers

ND or U	Indicates element was analyzed for, but not detected at or above the reporting limit.
J	Indicates a value greater than or equal to the method detection limit, but less than the quantitation limit.
H	Indicates analytical holding time exceedance.
B	Indicates that the analyte is found in the associated blank, as well as in the sample.
MSA	Indicates value determined by the Method of Standard Addition
E	This flag identifies compounds whose concentrations exceed the calibration range of the instrument for that specific analysis.
R	RPD outside accepted recovery limits
RL	Reporting limit; defined as the lowest concentration the laboratory can accurately quantitate.
S	Spike Recovery outside accepted recovery limits.
W	Post-digestion spike for Furnace AA analysis is out of control limits (85-115), while sample absorbance is less than 50% of spike absorbance.
*	Duplicate analysis not within control limits.
+	Indicates the correlation coefficient for the Method of Standard Addition is less than 0.995
#	See Case Narrative

Report Comments:

1. Soil, sediment and sludge sample results are reported on a "dry weight" basis.
2. Reporting limits are adjusted for sample size used, dilutions and moisture content, if applicable.

AMRO Environmental Laboratories Corp.

Date: 10-Jun-10

CLIENT: Ross Environmental Associates
Lab Order: 1006003
Project: 28-037 Hubbard Store
Lab ID: 1006003-01A

Client Sample ID: MW-1
Collection Date: 6/1/2010 9:53:00 AM
Matrix: AQUEOUS

Analyses	Result	RL	Qual	Units	DF	Date Analyzed
EPA 8260B AROMATIC VOLATILES BY GC/MS		SW8260B				Analyst: SK
Methyl tert-butyl ether	ND	1.0		µg/L	1	6/4/2010 1:47:00 PM
Benzene	ND	1.0		µg/L	1	6/4/2010 1:47:00 PM
Toluene	ND	1.0		µg/L	1	6/4/2010 1:47:00 PM
Ethylbenzene	ND	1.0		µg/L	1	6/4/2010 1:47:00 PM
m,p-Xylene	ND	2.0		µg/L	1	6/4/2010 1:47:00 PM
o-Xylene	ND	2.0		µg/L	1	6/4/2010 1:47:00 PM
1,3,5-Trimethylbenzene	ND	1.0		µg/L	1	6/4/2010 1:47:00 PM
1,2,4-Trimethylbenzene	ND	1.0		µg/L	1	6/4/2010 1:47:00 PM
Naphthalene	ND	1.0		µg/L	1	6/4/2010 1:47:00 PM
Surr: Dibromofluoromethane	97.9	82-122		%REC	1	6/4/2010 1:47:00 PM
Surr: 1,2-Dichloroethane-d4	95.8	73-135		%REC	1	6/4/2010 1:47:00 PM
Surr: Toluene-d8	97.3	82-117		%REC	1	6/4/2010 1:47:00 PM
Surr: 4-Bromofluorobenzene	93.6	77-119		%REC	1	6/4/2010 1:47:00 PM

AMRO Environmental Laboratories Corp.

Date: 10-Jun-10

CLIENT: Ross Environmental Associates
Lab Order: 1006003
Project: 28-037 Hubbard Store
Lab ID: 1006003-01A

Client Sample ID: MW-1
Collection Date: 6/1/2010 9:53:00 AM
Matrix: AQUEOUS

Analyses	Result	RL	Qual	Units	DF	Date Analyzed
GASOLINE RANGE ORGANICS (MOD. 8015B)		SW8260B				Analyst: SK
Gasoline Range Organics	ND	100		µg/L	1	6/2/2010 1:37:00 PM
Surr: Dibromofluoromethane	92.6	85-116		%REC	1	6/2/2010 1:37:00 PM
Surr: 1,2-Dichloroethane-d4	89.7	77-127		%REC	1	6/2/2010 1:37:00 PM
Surr: Toluene-d8	93.6	86-114		%REC	1	6/2/2010 1:37:00 PM
Surr: 4-Bromofluorobenzene	87.2	79-117		%REC	1	6/2/2010 1:37:00 PM
