AMALGAM SEPARATOR INSPECTION LOG Calendar Year 2007

Amalgam separators should be inspected weekly for drips or leaks. Any maintenance needed or performed should be noted. If the separator sludge collector or cartridge is made of clear plastic, note the level of sludge accumulation. Note the date you replace the collector box or cartridge and follow manufacturer instructions for recycling and obtaining a replacement collector box or cartridge. If you have any questions, you can contact the Dept. of Environmental Conservation at 800-974-9559.

Inspection Date	Initials	Leaks/Problems?	Maintenance Performed/Comments
1/3	GF	OK	None
1/10	DR	OK	Sludge level 2/3 full
1/17	GF	OK	Ordered new cartridge
1/25	GF	OK	None
2/2	GF	OK	None
2/9	GF	OK	Replaced cartridge
2/16	DR	OK	None

SAMPLE